

ACS GUIDELINE FOR CLINICAL FLOW CYTOMETRY LABORATORY PRACTICE

Second Edition 2017

Paper-based publications

This work is copyright. You may reproduce the whole or part of this work in unaltered form for your own personal use or, if you are part of an organisation, for internal use within your organisation, but only if you or your organisation do not use the reproduction for any commercial purpose and retain this copyright notice and all disclaimer notices as part of that reproduction. Apart from rights to use as permitted by the *Copyright Act 1968* or allowed by this copyright notice, all other rights are reserved and you are not allowed to reproduce the whole or any part of this work in any way (electronic or otherwise) without first being given the specific written permission from the Commonwealth to do so. Requests and inquiries concerning reproduction and rights are to be sent via e-mail to clinicalguidelines@cytometry.org.au

Internet sites

This work is copyright. You may download, display, print and reproduce the whole or part of this work in unaltered form for your own personal use or, if you are part of an organisation, for internal use within your organisation, but only if you or your organisation do not use the reproduction for any commercial purpose and retain this copyright notice and all disclaimer notices as part of that reproduction. Apart from rights to use as permitted by the *Copyright Act 1968* or allowed by this copyright notice, all other rights are reserved and you are not allowed to reproduce the whole or any part of this work in any way (electronic or otherwise) without first being given the specific written permission from the ACS.

Requests and inquiries concerning reproduction and rights are to be sent via email to: clinicalguidelines@cytometry.org.au

First published 2005

Second edition 2017 renamed and reformatted to be read '*ACS Guideline for Clinical Flow Cytometry Laboratory Practice*'.

CONTENTS

SCOPE	v
ABBREVIATIONS	vi
DEFINITIONS.....	vi
INTRODUCTION	1
1. PRE ANALYTICAL PHASE.....	2
1.1 Specimen Collection and Storage	2
1.2 Specimen Transport.....	2
1.3 Specimen Check.....	2
1.4 Antibody Reagents	3
2. ANALYTICAL PHASE	4
2.1 Sample analysis	4
.2 Performance Measures	4
3. POST ANALYTICAL PHASE	6
3.1 Reports	6
3.2 Data Storage	6
3.3 Laboratory Staff	7
REFERENCES CITED.....	8
NPAAC Reference Documents.....	8
Procedural References.....	9
Editorial Committee	9

The Australasian Cytometry Society (ACS) was established in 1979 and incorporated in 1992 with the aim of promoting research, development and applications in, and to disseminate knowledge of flow cytometry.

A function of the ACS is to assist with development and application of clinical flow cytometry applications for hospitals and laboratories in the diagnosis and treatment of disease. This includes the preparation of guidelines and education programs.

Guidelines produced by the ACS are issued as reference material to provide laboratories and accrediting agencies with minimum requirements for testing considered acceptable for good laboratory practice.

Failure to follow these guidelines may pose a risk to public health and patient safety.

SCOPE

The '*Guidelines for Clinical Flow Cytometry Laboratory Practice*' is an ACS document to be read in conjunction with the NPAAC Standards document '*Requirements for Medical Pathology Services*' or IANZ equivalent. The latter is the overarching document broadly outlining Standards for good medical pathology practice where the primary consideration is patient welfare, and where the needs and expectations of patients, Laboratory staff and referrers (both for pathology requests and inter-Laboratory referrals) are safely and satisfactorily met in a timely manner.

Whilst there must be adherence to all the Requirements in the NPAAC Standards document, reference to specific Standards in that document are provided for assistance under the headings in this document.

The NPPAC Requirements encompass:

- documentation and accurate patient identification systems to minimise clerical errors and misidentification
- retention of records, data and documentation
- requirements necessary for the use of computers and computer software in clinical flow cytometry Laboratory practice
- quality assurance and quality control programmes for reagents, techniques and personnel

The '*Guidelines for Clinical Flow Cytometry Laboratory Practice*' document is a consensus recommendation by ACS members and associates for best medical laboratory practice for a procedure, method, staffing resource or facility. It encompasses any measures, procedures or considerations unique to operation of a clinical flow cytometry laboratory where they differ from a routine clinical pathology laboratory. It is a guideline not a standard.

The '*Guidelines for Clinical Flow Cytometry Laboratory Practice*' document is for use by clinical flow cytometry Laboratory personnel for the purpose of clinical flow cytometry testing and safe laboratory practice. In house departmental laboratory procedural and quality control documentation eg Haematology or Immunology Departments should be read in conjunction with this ACS document.

For details on procedural matters and methodologies readers are directed to the Reference List at the end of this document for ACS recommended published peer reviewed texts, articles and guidelines. Definitions given above are sourced from NPAAC.

ABBREVIATIONS

FCS	means fluorescence correlation spectroscopy
IVD	means in vitro diagnostic
IANZ	means International Accreditation New Zealand, accreditation body of the Testing Laboratory Registration Council of New Zealand
NPAAC	means National Pathology Accreditation Advisory Council
RMPS	means Requirements for Medical Pathology Services, Tier 2, NPAAC Standard
7AAD	7 Aminoactinomycin D, a viability marker with DNA specificity

DEFINITIONS

CD	Cluster definition number used to identify individual markers eg CD3 for the pan T cell antigen
Cocktail	means an antibody reagent test mixture pre-prepared for use over the period of time validated.
Competent clinical flow cytometrist	means a person who has a minimum of two years clinical flow cytometry experience, and who has been documented to be competent in clinical flow cytometry according to the Laboratory's Quality System
External Quality Assessment	means a program in which multiple specimens are periodically sent to members of a group of Laboratories for analysis and/or identification, in which each Laboratory's results are compared with those of other Laboratories in the group and/or with an assigned value, and reported to the participating Laboratory and others. Such a program may also compare an individual's results with their peer group.
Flow test	means a test carried out by flow cytometry
Guideline	means a consensus recommendation for best medical laboratory practice for a procedure, method, staffing resource or facility
Guidelines for Clinical Flow Cytometry Laboratory Practice	means the overarching document broadly outlining standards for good clinical flow cytometry laboratory practice where the primary consideration is patient welfare, and where the needs and expectations of patients, Laboratory staff and referrers (both for pathology requests and inter-Laboratory referrals) are safely and satisfactorily

	met in a timely manner.
In vitro diagnostic medical device (IVD)	<p>means a medical device test if it is a reagent, calibrator, control material, kit, Specimen receptacle, software, instrument, apparatus, equipment or system, whether used alone or in combination with other diagnostic goods for in vitro use.</p> <p>It must be intended by the manufacturer to be used in vitro for the examination of Specimens derived from the human body, solely or principally for the purpose of giving information about a physiological or pathological state, a congenital abnormality or to determine safety and compatibility with a potential recipient, or to monitor therapeutic measures.</p> <p>The definition of an IVD does not encompass products that are intended for general Laboratory use that are not manufactured, sold or presented for use specifically as an IVD.</p>
In-house IVD	<p>means an IVD that is developed de novo, or developed or modified from a published source, or developed or modified from any other source, or its intended purpose, within the confines or scope of a Laboratory or Laboratory network, and is not supplied for use outside the Laboratory or Laboratory network.</p> <p>Commercial IVDs being used clinically for a purpose other than that originally intended by the manufacturer are also classed as in-house IVDs and are subject to the requirements of this standard.</p>
marker	means an antibody directed to an antigen of interest in or on a cell used for diagnostic purposes
Quality Assessment	means a measurement and monitoring function of quality assurance for determining how well health care is delivered in comparison with applicable standards or acceptable bounds of care.
Quality Assurance	means part of quality management focused on providing confidence that quality requirements will be fulfilled.
Quality control	<p>means the study of those errors that are the responsibility of the Laboratory, and the procedures used to recognise and minimise them. This study includes all errors arising within the Laboratory between the receipt of the Specimen and the dispatch of the report. On some occasions, the responsibility of the Laboratory may extend from the collection of the Specimen from the patient and the provision of a suitable container, to the dispatch and delivery of the report.</p> <p>Internal quality control:</p>

	means processes and activities that are used within the Laboratory to monitor the day-to-day operational and analytical performance of test procedures. These activities may include on-going instrument standardisation checks, instrument maintenance, analysis of control material, statistical or graphical assessment of results from control material.
Requirements for Medical Pathology Services (RMPS)	means the overarching document broadly outlining standards for good medical pathology practice where the primary consideration is patient welfare, and where the needs and expectations of patients, Laboratory staff and referrers (both for pathology requests and inter-Laboratory referrals) are safely and satisfactorily met in a timely manner. The standard headings are set out below – Standard 1 – Ethical Practice Standard 2 – Governance Standard 3 – Quality Management Standard 4 – Personnel Standard 5 – Facilities and Equipment A – Premises B – Equipment Standard 6 – Request-Test-Report Cycle A – Pre-Analytical B – Analytical C – Post-Analytical Standard 7 – Quality Assurance
stain	means bind monoclonal antibodies to markers on cells of interest to
standard	Means a minimum requirement for a procedure, method, staffing resource or laboratory facility that is required before a laboratory can attain accreditation

INTRODUCTION

This ACS document is intended to be used in clinical flow cytometry Laboratories to provide guidance on good practice in relation to flow cytometry and to assist assessors carrying out Laboratory accreditation assessments.

These Guidelines are intended to serve as consensus recommendations for best medical laboratory practice, have been developed by ACS members and associates with reference to other guidelines as published in peer reviewed journals. The ACS is the pre-eminent specialist society in clinical flow cytometry Laboratory practice in Australasia and New Zealand and has developed a number of guidelines relevant to this document.

These are Guidelines and not Standards. These Guidelines should be read in conjunction with the current version of the NPAAC Standards Tier 2 document '*Requirements for Medical Pathology Services*' or IANZ equivalent.

For clarification Standards are described:

- A Standard is the minimum requirement for a procedure, method, staffing resource or laboratory facility that is required before a laboratory can attain accreditation. The use of the verb 'must' in standards indicates mandatory requirements for pathology practice.

In each section of this document, points deemed important for practice are identified as either 'Guidelines' or 'Commentaries', as follows:

- A Guideline is a consensus recommendation for best medical laboratory practice for a procedure, method, staffing resource or facility. Guidelines are prefaced with a 'G' (e.g. G2.2). The use of the word 'should' in each Guideline within this document indicates a recommendation for good pathology practice.
- A Commentary may be provided to give clarification to the Guidelines as well as to provide examples and guidance on interpretation. Commentaries are prefaced with a 'C' (e.g. C1.2) and are placed where they add the most value.

Appendices if attached to this document are informative, that is explanatory in nature and may provide examples or information of a clinical nature and should be considered to be an integral part of this document.

Note: APC documents can be accessed at www.cytometry.org.au

1. PRE ANALYTICAL PHASE

Refer to NPAAC document '*Requirements for Medical Pathology Services*' for additional standards relating to minimum specimen labelling requirements, request forms, laboratory records, collection and transport conditions.

1.1 Specimen Collection and Storage

For flow cytometry analysis appropriate anticoagulant and storage be used according to sample type and disease investigation^(1; 2). Refer to assay specific ACS guidelines for further information.

1.2 Specimen Transport

Refer to assay specific ACS test guidelines for transport, storage and rejection criteria conditions.

G1.2 Specimen Viability

Non-viable cells are significant source of false positive staining. Viability testing is recommended for some tests on samples more than 24 hours after collection or if there is obvious deterioration of the sample^(2; 3).

C1.2 Viability can be tested using 7AAD^(2; 4) or tryptophan blue.
Propidium iodide is not recommended due to instrument contamination and carryover.

1.3 Specimen Check

G1.3.1 Primary collection samples should have the patients name and two identifiers (eg date of birth, medical record number, test/accession number), including collection date, sample type and/or collection site.

C1.3.1 Secondary sample/assay tubes should have patient name or part thereof and at least one identifier. Barcode alone is not acceptable.

G1.3.2 A protocol for return unlabeled/mismatch specimens according and criteria for rejection should be followed as described in the laboratory department manual.

1.4 Antibody Reagents

G1.4.1 New antibodies introduced into the laboratory should be correctly validated by clinical correlation ⁽²⁾.

C1.4.1(i) Antibodies should be tested on a commercial reagent control or appropriate sample where available to determine whether they can be used to measure the level of antigen.

C1.4.1(ii) New Lot number and delivery of reagents should be tested where control reagents or suitable blood samples are available.

C1.4.1(iii) New Lot numbers, delivery of reagents, expiry dates and in use dates should be recorded for each reagent.

G1.4.2 Use of reagents beyond expiry dates is not recommended.

C1.4.2(i) If used beyond expiry date they should be validated against in-date reagents, and with a positive control tested at time of use.

C1.4.2(ii) It should be noted reagent fluorescence intensity may vary significantly beyond reagent expiry.

G1.4.3 Cocktail antibody preparations for flow tests should be correctly validated and each lot tested prior to use ⁽¹⁾.

G1.4.4 Reagent titration is acceptable where manufacturer’s instructions allow and following IVD guidelines.

C1.4.4(i) Repeat testing is required for new clones, manufacturers and catalog numbers.

C1.4.4(ii) Excess antibody may lead to increased non-specific binding and quenching of fluorescence signals ⁽³⁾.

2. ANALYTICAL PHASE

Refer to NPAAC document ‘*Requirements for Medical Pathology Services*’ for additional standards relating to sample analysis and performance measures.

2.1 Sample analysis

G2.1.1 Sample preparation for any assay should seek to minimize manipulation of the cells and maximize preservation of their viability and antigenic integrity.

G2.1.2 Appropriate software and gating should be used for individual flow tests ^(2; 5).

.2 Performance Measures

G2.2.1 Analysers should be set up, monitored, maintained and documented using appropriate material, reviewed by a competent clinical flow cytometrist ^(2; 1).

C2.2.1(i) Individual assay settings should be monitored and reanalyzed as need e.g. daily QC beads for analyser performance, assay settings ^(3; 2) .

C2.2.1(ii) Variation from manufacturer's advice for use of instrumentation and reagents needs to follow current recommendations for IVD medical device and In-house IVDs.

G2.2.2 A control should be prepared and run on a regular basis in parallel with patient samples where appropriate. A positive (and/or negative) reagent control should be run either daily or with each assay performed.

C2.2.2 For some assays eg Lymphoma and Leukaemia screening, it is not practical or necessary to analyse a normal control sample on a daily or weekly basis if the laboratory is active and within-run positive and negative control results demonstrate appropriate reactivity.

G2.2.3 Where absolute numbers (eg cells/uL) are reported, a control reagent should have specified ranges for the analytes measured, and reasons for and deviations determined.

G2.2.4 A minimum number of target cellular events should be collected according to the flow test, reporting style suited (quantitative or qualitative) and criteria for rejection determined for individual flow tests.

G2.2.5 Sensitivities for each assay should be calculated, documented and reviewed when new assays are setup or when alterations to assay methods, technologies or reagents change ⁽¹⁾.

G2.2.6 Each laboratory should establish reference limits for antigens being tested where appropriate.

3. POST ANALYTICAL PHASE

Refer to NPAAC document *'Requirements for Medical Pathology Services'* for additional standards relating to reports, record keeping, result validation.

3.1 Reports

G3.1.1 Interpretation and report comments should be made and verified by competent clinical flow cytometrist or suitably trained medical officer.

G3.1.2 Where quantitative (numerical) results are given, reference ranges should be provided where appropriate ⁽²⁾.

G3.1.3 Reports should be completed in a timely manner and in a time commensurate with clinical need. Depending on the test this should be less than a few hours eg CD34 pre harvest counts to no longer than 5 working days for haematology oncology reports.

C3.1.3 Adequate trained staff and resources must be provided to meet these targets.

3.2 Data Storage

G3.2 Electronic data (FCS) should have backup copies and storage for the period required by NPAAC, IANZ.

C3.2(i) FCS data should be traceable and include sample identifiers such as patient details, date and sample type ⁽³⁾.

C3.2 (ii) FCS files should be traceable, matching correctly to individual samples from a patient ⁽²⁾.

C3.2 (iii) use of barcoding and reference to worklist data may help reduce error.

3.3 Laboratory Staff

G3.3.1 Staff in a flow cytometry facility need a high level of skill to perform a unique variety of clinical tests using specialised equipment and software. There should be documented evidence staff are given sufficient time and resources to achieve and maintain their skills⁽²⁾.

C3.3.1(i) A competent flow cytometrist able to work unsupervised at all levels should have a minimum of two years clinical flow cytometry experience, and have documented competency in clinical flow cytometry according to the Laboratory's Quality System⁽³⁾.

C3.3.1 (ii) A sufficient number of cases studied over a given period is required to maintain proficiency⁽¹⁾. This includes laboratory and medical staff.

C3.3.1 (iii) Staff reporting results should be annually tested for proficiency in use of gating software and interpretation of results.

G3.3.1 Clinical flow cytometry laboratories should be supervised by a scientist with demonstrated experience in all aspects of flow cytometry testing, interpretation and quality management.

REFERENCES CITED

1. Davis B, Wood B, et al, 'Validation of Cell-Based Fluorescence Assays: Practice Guidelines from the ICSH and ICCS' *Cytometry Part B* 84B:282-285, 2013
2. H42-A2 Vol. 27 No. 16 Enumeration of Immunologically Defined Cell Populations by Flow Cytometry; Approved Guidelines – Second Edition. Clinical and Laboratory Standards Institute, 2007
3. H43 – A2 Vol. 27 No. 11 “Clinical Flow Cytometric Analysis of Neoplastic Hematolymphoid Cells; Approved Guideline – Second Edition. Clinical and Laboratory Standards Institute, 2007
4. Shenkin M, Babu R, Maise R, 'Accurate Assessment of Cell Count and Viability With a Flow Cytometer', *Cytometry Part B (Clinical Cytometry)*, 72B:427-432, 2007
5. Johansson U. et al, 2014, 'Guidelines on the use of multicolour flow cytometry in the diagnosis of haematological neoplasms', *British Journal of Haematology*, John Wiley & Sons Ltd
6. Brent Wood et al, '2006 Bethesda International Consensus Recommendations on the Immunophenotypic Analysis of Hematolymphoid Neoplasia by Flow Cytometry: Optimal Reagents and Reporting for the Flow Cytometric Diagnosis of Hematopoietic Neoplasia', *Cytometry Part B (Clinical Cytometry)* 72B:S14–S22, 2007

NPAAC Reference Documents

Requirements for Medical Pathology Services, First Edition, 2013

Print ISBN: 978-1-74241-913-8, Online ISBN: 978-1-74241-914-5, Publications approval number: 10207

Regulatory Requirements for In-house IVDs, Version 2.0, March 2016, Australian Government, Department of Health, Therapeutic Goods Administration

Requirements for the Development and Use of In-house In Vitro Diagnostic Devices (IVDs), Third Edition 2014

ISBN: 1 74186 158 6; Online ISBN: 1 74186 159 4; Publications Approval Number: 3957

PROCEDURAL REFERENCES

Further detail outside the scope of this document can be found in the references listed above to assist with further detailed descriptions of setting up and maintaining flow cytometry assays, instrumentation and quality systems.

Editorial Committee

Sheree Bailey, Neil Came, Kate Marson, Neil McNamara, Vuong Nguyen, Michelle Petrasich, William Sewell, Sandy Smith, Elizabeth Todorova, Susan Wright

Further ACS clinical flow cytometry guidelines documents are available on the website:

www.cytometry.org.au

Email: clinicalguidelines@cytometry.org.au